

GAMMEX

MAMMOGRAPHY



Stereotactic Breast Biopsy Phantom

GAMMEX 164A

The phantom provides a good representation of breast tissue, thereby making it a close to real-life learning experience.

The automated stereotactic breast biopsy procedure depends on several variables for accurate needle placement. Use of the Gammex 164A provides technology and physicians with the training needed to be confident in performing stereotactic procedures.

The phantom is made of clear gel encased in a soft vinyl for easy compression and a skin-like resistance to needle insertion. Embedded in the gel are 20 to 25 radiopaque lesions ranging in size from 2 to 5 mm. The 3 and 5 mm gel lesions are used for practicing core

biopsies and the 2 mm liquid lesions allow for the practice of fine needle aspiration and test the accuracy of the biopsy system and the operator.

SPECIFICATIONS

Construction Gel with attenuation properties similar to breast tissue.
Outer casing Vinyl
Radiopaque
Lesions 2 to 5 mm
Solid Gel for core biopsy
Liquid dye for fine needle aspiration
Compressible within biopsy instrument

Gammex

WWW.GAMMEX.COM

PEO B.V.

info@gotopeo.com
www.gotopeo.com

The Netherlands

Havenweg 16, 6603 AS Wijchen
+31 (0)24 648 86 88

Belgium

Watermolenstraat 2, B-2910 Essen
+32 (0)3 309 32 09

CoC 34107894

VAT NL807859151B01

IBAN NL29 RABO 0356 1960 46

BIC RABONL2U

